

SENSOTEC

2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850 - 5000
INTERNET URL <http://www.sensotec.com>

CERTIFICATE OF CALIBRATION

MODEL: TH/1590-05
ORDER CODE: BL911MX
SERIAL NUMBER: 967551
CALIBRATION DATE: Aug 21/2003
INPUT RESISTANCE: $352.\Omega$
OUTPUT RESISTANCE: $352.\Omega$
LEAKAGE: ∞


WIRING CODE

CAPACITY: 500 KN
COMPRESSION
CALIBRATED AT: 500 KN
EXCITATION: 10.0 VDC
CALIBRATION FACTOR: 2.004 MV/V
SHUNT RESISTOR: 59k Ω
SHUNT CAL FACTOR: 1.488 MV/V

UNAMP#1,4-COND,CBL

COLOR	DESIGNATION
RED	(+)EXCITATION
BLACK	(-)EXCITATION
GREEN	(-)OUTPUT
WHITE	(+)OUTPUT

001-0333-01



Accepted and Certified by:

Michael A Stanley

Date Printed: 8/22/2003