

# Honeywell


2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850 - 5000

## CERTIFICATE OF CALIBRATION

MODEL: JEC-AG  
PART NO.: 060-H032-03  
ORDER CODE: AY322HR  
SERIAL NUMBER: L12451100  
CALIBRATION DATE: 02/27/2009

RANGE: +/-2.000 INCHES  
LINEARITY: 0.1600%  
SUPPLY: 40.0 VDC  
CALIBRATION FACTOR: 2.4860 V/INCH

### WIRING CODE



Accepted and Certified by:

*Eszter Gozon*

Date Printed: 3/11/2009

Eszter Gozon, Quality Manager