

Honeywell

2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850-5000




CERTIFICATE OF CALIBRATION

MODEL: JEC-AG
 PART NO.: 060-H032-03
 ORDER CODE: AY322HR
 SERIAL NUMBER: L11624100
 CALIBRATION DATE: 7/16/2008

RANGE: +/- 2.0000 INCHES
 LINEARITY: 0.1100 %
 SUPPLY: 40.0000 VDC
 CALIBRATION FACTOR: 2.4580 V/INCH

WIRING CODE



Accepted and Certified by:

Eszter Gozon

Date Printed: 7/16/2008

Eszter Gozon, Quality Manager