

CORRIGENDUM

Date: 20/01/2026

Tender ID: 2026_IITK_894526_1
Tender Title: Supply, Installation, and Commissioning of Electron Microscope
(Dual Bean FIBSEM System with Accessories)
Tender Ref. No. PI/ACMS/RD/2025-26/NPG/NC-10

(1)

The following missed out part of the technical specification at page no.38 of the tender document is required to be considered for the Supply, Installation, and Commissioning of Electron Microscope (Dual Bean FIBSEM System with Accessories):

3. Optional Accessory: Carbon Coater / Evaporator

S. No.	Specifications
3.1	Carbon Coater <ul style="list-style-type: none">○ Chamber size: should be at least 150 mm Ø diameter, 210 mm height○ Ultimate vacuum should be at least 6×10^{-3} mbar○ Operational vacuum should be at least 1×10^{-1} mbar○ User interface Touch screen color display is essential○ Evaporation current (rod) should be in the range of 5-75 A○ Carbon source should be minimum 3.2 mm rods or carbon fibre.○ Outgas current range should be in the range of 1-40 A

(2)


Terms and Conditions:

General terms and condition	6. Vendor shall quote for the system with required hardware, software, spares, consumables and accessories. Separate technical and financial bids should be provided, and all the technical details should be provided in the Technical Bid.
-----------------------------	--

(3)

On page no. 35 & serial no. 1.30. Manipulator. There is a typographical error in the sub section which mention the point as 3.34.1 whereas it should be 1.30.1.

All bidders are requested to consider the above-mentioned incorporated changes made in the specifications in the technical bid and quote accordingly and provide a sperate price bifurcation for the optional accessories. Thank You



Dr. Nilesh Prakash Gurao
Indenter, EM (DB-FIBSEM)
ACMS, IITK