

INDIAN INSTITUTE OF TECHNOLOGY KANPUR
DEPARTMENT OF CIVIL ENGINEERING

Enquiry No: CE/HWRE/SG/2015-16/1

Date: Aug 13, 2015

Subject: Quotation for supply of ANALYTICAL BALANCE :- (BRAND: SARTORIUS, METTLER-TOLEDO)

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us by Sep 04, 2015 in the form of a hard copy to the address mentioned below. If you have any questions please call Dr. S. Guha at 0512-2597917, email:sguha@iitk.ac.in.

The prospective suppliers are required to send quotation in two parts in sealed envelopes, as “Technical Bid” and “Financial Bid”. The Technical Bid should contain detailed technical specification of the product being offered and should not mention any prices. The Financial Bid should include the detailed price quotation clearly including the cost of the equipment, taxes, service charges, shipping and handling charges, if any. The two separate and sealed envelopes should be clearly marked appropriately as “Technical Bid” and “Financial Bid”. Kindly write the inquiry no. on the top of envelop.

Specification for Analytical Balance :- (Brand: Sartorius, Mettler-Toledo)

Maximum capacity	60-120 g
Minimum readability	0.01 mg
Repeatability	≤ 0.08 mg
Linearity	≤ ±0.015 mg
Settling Time	Approx. 8 s
Connectivity	RS232 to connect to PC, Bluetooth to connect to PC
Calibration	Fully automatic time and temperature controlled internal adjustment
Size of weighing pan	80-85 mm diameter
Regulatory requirements	ISO or any other relevant trademark
Operation	Keys for main basic functions
Adjustment	Internal, isoCal
Overload protection	Must Have

Terms and Conditions

1. Warranty/Guarantee should be clearly mentioned
2. Maximum education discount, if any should be offered.
3. Validity of quotation should be at least for 60 days.
4. Prices should be quoted on FOB (Free on Board) only (if imported).
5. Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/inspection).
6. Quotation should carry proper certification like agency certificate, proprietary certificate, etc.

Kindly write the enquiry number on the top of envelop.

Dr. Saumyen Guha
Department of Civil Engineering
IIT Kanpur, Kanpur 208016