

Request for Quotation

REF: AE/AK/PG/03

Sub : Quotation for supply of : **Continuous emission monitoring system**

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us before 15/2/2012 in the form of hard copy / soft copy.

The soft copy of the model quotation can also be downloaded from our website

Technical details of the store are as follows;

Measurement of CO, CO₂, O₂, NO_x, THC and VOC

CO, CO₂, O₂, Non Dispersive Infra Red

Measurement ranges:

CO: 10 to 50,000 ppm,

CO₂: 100 to 250,000 ppm,

O₂: 0 to 25%v

NO_x measurement : Chemiluminescence

NO, NO_x and NO₂

Ranges: 0-10000 ppm

THC and VOC : Heated FID

Ranges: 0 – 10000 ppm.

Best regards,
Sincerely,

A. Kushari
Department of Aerospace Engineering
IIT Kanpur, India.